

CBC-3D[®]

For Zybio
HEMATOLOGY CONTROLS

CONTROL

Assay values and ranges provided by Zybio

LOT B1124



reference tag

Instrument Model	Parameter	Low	Normal	High	
		LOT B1124L	LOT B1124N	LOT B1124H	
Z3,Z3 CRP	WBC	10 ⁹ /L	2.10 ± 0.42	8.05 ± 1.21	22.14 ± 3.32
	Lym#	10 ⁹ /L	1.25 ± 0.25	2.65 ± 0.53	3.98 ± 1.59
	Mid#	10 ⁹ /L	0.17 ± 0.14	0.49 ± 0.39	0.93 ± 0.74
	Gran#	10 ⁹ /L	0.68 ± 0.20	4.92 ± 0.74	17.24 ± 1.72
	Lym%	%	59.5 ± 11.9	32.9 ± 13.1	18.0 ± 7.2
	Mid%	%	8.0 ± 6.4	6.0 ± 4.8	4.2 ± 3.3
	Gran%	%	32.4 ± 9.7	61.1 ± 12.2	77.8 ± 7.8
	RBC	10 ¹² /L	2.30 ± 0.14	4.51 ± 0.27	5.58 ± 0.33
	HGB	g/L	59 ± 4	136 ± 7	192 ± 8
	HCT	%	17.9 ± 1.6	39.3 ± 3.5	53.7 ± 4.8
Z30,Z30 CRP	MCV	fL	77.9 ± 5.5	87.3 ± 6.1	96.3 ± 6.7
Z31,Z31 CRP	MCH	pg	25.6 ± 1.8	30.1 ± 2.1	34.4 ± 2.4
	MCHC	g/L	329 ± 26	344 ± 28	358 ± 29
	RDW-CV	%	17.3 ± 3.0	16.5 ± 3.0	15.3 ± 3.0
	RDW-SD	fL	46.7 ± 6.0	49.7 ± 6.0	50.2 ± 6.0
	PLT	10 ⁹ /L	67 ± 13	269 ± 40	554 ± 83
	MPV	fL	9.8 ± 3.0	9.9 ± 3.0	9.8 ± 3.0
	PDW	fL	16.4 ± 3.0	16.3 ± 3.0	16.4 ± 3.0
	PCT	%	0.065 ± 0.050	0.266 ± 0.100	0.545 ± 0.200
P-LCC	10 ⁹ /L	17 ± 10	67 ± 22	136 ± 30	
P-LCR	%	25.8 ± 9.0	24.7 ± 9.0	24.6 ± 9.0	

Notes: Before using, refer to the instruction sheet for mixing directions.
CBC-3D for Zybio has 14-day open vial stability.

Zybio Inc.

Add: Floor 1 to Floor 5, Building 30, No.6 of Taikang Road, Block C of Jianqiao Industrial Park, Dadukou District, Chongqing, China
Tel: +86(0)23 6895 9999 E-mail: info@zybio.com Web: www.zybio.com