CBC-3D[®]

for Abacus HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT B1124

QCP DATA MONTHS: NOVEMBER, DECEMBER, JANUARY


2025-02-05

Instrument	Parameter	Low	Normal	High
		LOT B1124L	LOT B1124N	LOT B1124H
Abacus 3CP	WBC \times 10 9 /L	2.0 ± 0.5	7.5 ± 1.2	20.7 ± 2.4
	$LYM \times 10^9/L$	1.2 ± 0.3	2.9 ± 0.6	4.0 ± 1.5
	MID × 10 ⁹ /L	0.2 ± 0.2	0.4 ± 0.4	1.0 ± 1.0
	GRA × 10 ⁹ /L	0.6 ± 0.2	4.3 ± 0.6	15.6 ± 1.7
	LYM %	61.5 ± 11.0	38.0 ± 7.0	19.5 ± 7.0
	MID %	10.0 ± 10.0	5.0 ± 5.0	5.0 ± 5.0
	GRA %	28.5 ± 9.0	57.0 ± 8.0	75.5 ± 8.0
	RBC × 10 ¹² /L	2.39 ± 0.25	4.78 ± 0.30	5.90 ± 0.35
	HGB g/dL	6.4 ± 0.6	13.9 ± 0.8	18.8 ± 1.2
	HCT %	19.1 ± 2.0	42.5 ± 2.4	57.2 ± 3.0
	MCV fL	80.0 ± 5.0	89.0 ± 5.0	97.0 ± 5.0
	MCH pg	26.8 ± 2.4	29.1 ± 2.8	31.9 ± 2.8
	MCHC g/dL	33.5 ± 3.0	32.7 ± 3.0	32.8 ± 3.0
	RDW CV %	17.5 ± 5.0	16.0 ± 5.0	15.5 ± 5.0
	PLT × 10 ⁹ /L	65 ± 30	241 ± 50	483 ± 60
	MPV fL	9.0 ± 3.0	8.8 ± 3.0	9.0 ± 3.0

Before using, refer to the instruction sheet for mixing directions.

For technical assistance in the USA and Canada call Technical Service at (800) 523-3395.

All brands and products are trademarks or registered trademarks of their respective companies.



R&D Systems, Inc. 614 McKinley Place NE Minneapolis, MN USA 55413 AS153-002 Rev. 04/14