

ADVIA[®] Retic Plus

HEMATOLOGY CONTROLS

CONTROL


+

ASSAY VALUES AND EXPECTED RANGES

QCP DATA MONTHS: **OCTOBER, NOVEMBER**


ADVIA 120, 2120 and 2120i LEVEL 1

LOT	AD1401	
Parameter	Target	Range
Reticulocyte %	1.8	± 1.0
Retic RBC × 10 ⁶ /μL	3.55	± 0.30
MCVg fL	67.0	± 7.0
MCVr fL	81.0	± 9.0

	2024-12-05	
Parameter	Target	Range
CHCMg g/dL	39.5	± 4.0
CHCMr g/dL	36.0	± 4.0
CHg pg	27.0	± 3.0
CHr pg	30.0	± 4.0

ADVIA 120, 2120 and 2120i LEVEL 2

LOT	AD1402	
Parameter	Target	Range
Reticulocyte %	5.7	± 2.4
Retic RBC × 10 ⁶ /μL	3.51	± 0.30
MCVg fL	67.5	± 7.0
MCVr fL	80.5	± 9.0

	2024-12-05	
Parameter	Target	Range
CHCMg g/dL	39.5	± 4.0
CHCMr g/dL	35.5	± 4.0
CHg pg	27.0	± 3.0
CHr pg	28.5	± 4.0

Before using, refer to the instruction sheet for mixing directions.

Upon scanning the barcodes, the filenames will default to CBC/Diff Low, Normal and High controls.

Rename the files to indicate that they are Retic Controls.

For technical assistance in the USA and Canada call Technical Service at (800) 523-3395.

All brands and products are trademarks or registered trademarks of their respective companies.

ADVIA[®] Retic Plus

HEMATOLOGY CONTROLS

CONTROL

++

ASSAY VALUES AND EXPECTED RANGES

QCP DATA MONTHS: **OCTOBER, NOVEMBER**

ADVIA 120, 2120 and 2120i LEVEL 3

LOT AD1403

Parameter	Target	Range
Reticulocyte %	9.8	± 3.5
Retic RBC × 10 ⁶ /μL	3.53	± 0.30
MCVg fL	68.5	± 7.0
MCVr fL	80.5	± 9.0



2024-12-05

Parameter	Target	Range
CHCMg g/dL	39.5	± 4.0
CHCMr g/dL	35.5	± 4.0
CHg pg	27.0	± 3.0
CHr pg	28.0	± 4.0

Before using, refer to the instruction sheet for mixing directions.

Upon scanning the barcodes, the filenames will default to CBC/Diff Low, Normal and High controls.

Rename the files to indicate that they are Retic Controls.

For technical assistance in the USA and Canada call Technical Service at (800) 523-3395.

All brands and products are trademarks or registered trademarks of their respective companies.