



## CONTROL FOR RETICULOCYTE CHECK ASSAY SHEET

Applicable model: BF-7200Plus/BF-7200

LOT: 3511100

Exp. date: 2026-02-01

AS737-001 Rev.02/25

| Parameter |             | Level: 1 |        |            | Level: 2 |        |                 | Level: 3 |        |                 |
|-----------|-------------|----------|--------|------------|----------|--------|-----------------|----------|--------|-----------------|
|           |             | Mean     | Limit  | Range      | Mean     | Limit  | Range           | Mean     | Limit  | Range           |
| RET%      | %           | 1.00     | 1.00   | 0 - 2.00   | 3.05     | 3.00   | 0.05 - 6.05     | 11.30    | 5.00   | 6.30 - 16.30    |
| RET#      | $10^{12}/L$ | 0.0300   | 0.0300 | 0 - 0.0600 | 0.1300   | 0.1300 | 0.0000 - 0.2600 | 0.2180   | 0.1300 | 0.0880 - 0.3480 |

二维码

level: 1

二维码

level: 2

二维码

level: 3



DIRUI INDUSTRIAL CO.,LTD.

95 Yunhe Street, New & High Tech.

Development Zone Changchun, Jilin 130012, P. R. China

Tel: +86 (431) 85100409

Fax: +86 (431) 85172581

E-mail: dirui@dirui.com.cn

<http://www.dirui.com.cn>