

Body Fluid-I


HEMATOLOGY CONTROLS

ASSAY VALUES AND EXPECTED RANGES

CONTROL

maccura®F800 (manual sampling)

QCP DATA MONTHS: NOVEMBER, DECEMBER

 2025-01-05

Parameters	LEVEL 1 LOT BFI1124-1			LEVEL 2 LOT BFI1124-2			LEVEL 3 LOT BFI1124-3		
	Assay Mean	±Limit	Expected Range	Assay Mean	±Limit	Expected Range	Assay Mean	±Limit	Expected Range
WBC-BF (10 ⁹ /L)	0.024±0.019		0.005~0.043	0.196±0.069		0.127~0.265	0.898±0.225		0.673~1.123
RBC-BF (10 ¹² /L)	0.024±0.019		0.005~0.043	0.064±0.022		0.042~0.086	0.485±0.121		0.364~0.606

- NOTES:**
1. Before using, confirm the software for the latest version, refer to the instruction sheet for mixing directions, and choice the QC mode.
 2. All brands and products are trademarks or registered trademarks of their respective companies, All results are for reference only.

Body Fluid-I


HEMATOLOGY CONTROLS

ASSAY VALUES AND EXPECTED RANGES

CONTROL

maccura®F810 (manual sampling)

QCP DATA MONTHS: **NOVEMBER, DECEMBER**

 2025-01-05

Parameters	LEVEL 1 <input type="text" value="LOT"/> BFI1124-1			LEVEL 2 <input type="text" value="LOT"/> BFI1124-2			LEVEL 3 <input type="text" value="LOT"/> BFI1124-3		
	Assay Mean	±Limit	Expected Range	Assay Mean	±Limit	Expected Range	Assay Mean	±Limit	Expected Range
WBC-BF (10 ⁹ /L)	0.024±0.019		0.005~0.043	0.196±0.069		0.127~0.265	0.898±0.225		0.673~1.123
RBC-BF (10 ¹² /L)	0.024±0.019		0.005~0.043	0.064±0.022		0.042~0.086	0.485±0.121		0.364~0.606

- NOTES:**
1. Before using, confirm the software for the latest version, refer to the instruction sheet for mixing directions, and choice the QC mode.
 2. All brands and products are trademarks or registered trademarks of their respective companies, All results are for reference only.

Body Fluid-I


HEMATOLOGY CONTROLS

ASSAY VALUES AND EXPECTED RANGES

CONTROL

maccura®F880 (manual sampling)

QCP DATA MONTHS: NOVEMBER, DECEMBER

 2025-01-05

Parameters	LEVEL 1 <input type="text" value="LOT"/> BFI1124-1			LEVEL 2 <input type="text" value="LOT"/> BFI1124-2			LEVEL 3 <input type="text" value="LOT"/> BFI1124-3		
	Assay Mean	±Limit	Expected Range	Assay Mean	±Limit	Expected Range	Assay Mean	±Limit	Expected Range
WBC-BF (10 ⁹ /L)	0.024±0.019		0.005~0.043	0.196±0.069		0.127~0.265	0.898±0.225		0.673~1.123
RBC-BF (10 ¹² /L)	0.024±0.019		0.005~0.043	0.064±0.022		0.042~0.086	0.485±0.121		0.364~0.606

- NOTES:**
1. Before using, confirm the software for the latest version, refer to the instruction sheet for mixing directions, and choice the QC mode.
 2. All brands and products are trademarks or registered trademarks of their respective companies, All results are for reference only.