


# CBC-CAL

For Getein Biotech  
HEMATOLOGY CALIBRATOR

CALIBRATOR

LOT PLUS0526

 2026-06-05

Instrument	Parameter	Assigned Value	Acceptable Range
BHA-5000 BHA-5100 BHA-5010 BHA-5110	WBC $10^3/\mu\text{L}$	10.02	9.82-10.22
	RBC $10^6/\mu\text{L}$	4.44	4.34-4.54
	HGB g/dL	13.3	13.1-13.5
	MCV fL	88.9	86.9-90.9
	PLT $10^3/\mu\text{L}$	263	251-275
BHA-3000 BHA-3100	WBC $10^3/\mu\text{L}$	8.51	8.31-8.71
	RBC $10^6/\mu\text{L}$	4.51	4.41-4.61
	HGB g/dL	13.5	13.3-13.7
	MCV fL	83.8	81.8-85.8
	PLT $10^3/\mu\text{L}$	268	256-280