

CBC-ST PLUS

For Getein Biotech
HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT ST1225

 2026-03-05

Instrument	Paramenter	Low	Normal	High
		LOT ST1225L	LOT ST1225N	LOT ST1225H
BHA-3000 BHA-3100 QC Mode	WBC 10 ⁹ /L	2.78±0.34	7.58±0.80	20.16±1.20
	Neu 10 ⁹ /L	1.76±0.22	3.83±0.91	7.47±1.16
	Lym 10 ⁹ /L	0.36±0.28	2.25±0.76	11.93±1.26
	Mxd 10 ⁹ /L	0.65±0.10	1.50±0.53	0.73±0.34
	Neu %	63.7±6.5	50.5±12.0	36.9±4.8
	Lym %	12.8±8.3	29.7±10.0	59.3±4.7
	Mxd %	23.5±2.8	19.8±7.0	3.9±2.0
	RBC 10 ¹² /L	2.53±0.24	4.02±0.22	5.83±0.36
	HGB g/L	63±4	113±5	179±16
	HCT %	17.1±1.7	33.1±2.4	3.9±2
	MCV fL	67.7±9.0	82.4±5.0	88.8±12.1
	MCH pg	24.8±3.0	28.1±2.5	30.7±2.3
	MCHC g/L	367±78	341±30	347±56
	RDW-CV %	14.3±3.0	13.8±3.0	14.2±3.8
	RDW-SD fL	43.5±4.1	43.3±8.0	50.4±7.8
	PLT 10 ⁹ /L	68±20	189±30	328±91
	MPV fL	9.5±1.9	8.5±3.0	8.6±0.4
	PDW	12.3±3.3	12.3±3.0	12.1±2.4
	PCT mL/L	0.65±0.27	1.61±1.00	2.82±0.77
	P-LCR %	22.5±11.7	20.4±10.2	21.5±10.5
	P-LCC 10 ⁹ /L	15±9	35±11	70±19