

HQ-6D

HEMATOLOGY CONTROLS

CONTROL

Assay values and ranges provided by URIT Medical

LOT**2504000****2025/05/25**

Instrument	Parameter	Low		Normal		High	
		LOT	25040001	LOT	25040002	LOT	25040003
BH 6100/BH 6180 BH 6200/BH 6280 Assay values established in QC file.	WBC	3.73	±0.80	7.33	±1.00	19.73	±2.50
	LYM%	10.22	±9.50	18.73	±10.00	16.91	±9.00
	LYM#	0.38	±0.38	1.37	±0.80	3.34	±1.60
	MON%	5.30	±5.30	9.39	±7.00	5.53	±5.53
	MON#	0.20	±0.20	0.69	±0.69	1.09	±1.09
	NEU%	69.47	±12.00	63.33	±10.00	70.63	±10.00
	NEU#	2.59	±0.50	4.64	±0.80	13.94	±2.00
	EOS%	14.42	±7.00	7.79	±6.50	4.88	±4.88
	EOS#	0.54	±0.54	0.57	±0.57	0.96	±0.96
	BASO%	0.59	±0.59	0.76	±0.76	2.05	±2.05
	BASO#	0.02	±0.02	0.06	±0.06	0.40	±0.40
	RBC	2.39	±0.18	4.65	±0.24	5.66	±0.30
	HGB	63	±4	135	±6	176	±8
	HCT	19.6	±2.0	41.3	±3.0	53.3	±4.0
	MCV	82.0	±5.0	88.8	±5.0	94.2	±6.0
	MCH	26.4	±2.5	29.0	±2.5	31.1	±2.5
	MCHC	321	±30	327	±30	330	±30
	RDW_CV	16.9	±5.0	17.8	±5.0	16.7	±6.0
	RDW_SD	44.0	±10.0	48.2	±10.0	48.4	±12.0
	NRBC%*	0.01	±0.01	0.01	±0.01	0.01	±0.01
PLT	44	±20	217	±45	407	±65	
MPV	8.6	±3.0	8.7	±3.0	8.9	±3.0	
PDW	10.7	±5.0	11.1	±5.0	11.4	±5.0	
PCT	0.04	±0.04	0.19	±0.10	0.36	±0.20	
P_LCR	22.8	±10.0	21.6	±10.0	23.0	±10.0	
P_LCC	10	±10	47	±25	94	±50	
RET%	0.63	±0.63	0.76	±0.76	0.76	±0.76	

Before using, refer to the instruction sheet for mixing directions.

URIT Medical Electronic Co., Ltd.

No. D-07 Information Industry District, High-Tech Zone, Guilin, Guangxi 541004, P.R.China.

Tel: +86(773)2288586 Fax: +86(773)2288560

Web: www.urit.com Email: service@uritest.com

EC-REPRESENTATIVE:

Shanghai International Holding Corp. GmbH (Europe)

Eiffestrasse 80, 20537 Hamburg, Germany

Manufactured for Shenzhen URIT under contract

AS719-000 Rev. 10/23