


CBC-3D[®]

for MINDRAY
HEMATOLOGY CONTROLS

CONTROL

ASSAY VALUES AND EXPECTED RANGES

LOT	B0516
	2016-08-05

Instrument	Parameter	Low	Normal	High
		LOT B0516L	LOT B0516N	LOT B0516H
MINDRAY	WBC 10 ³ /uL	2.0 ± 0.5	8.3 ± 1.0	21.4 ± 2.5
BC-2800Vet, BC-2600Vet	RBC 10 ⁶ /uL	2.39 ± 0.18	4.62 ± 0.24	5.88 ± 0.30
	HGB g/dL	6.1 ± 0.4	13.5 ± 0.6	19.1 ± 0.8
(Software version 2.0 or higher)	HCT %	18.0 ± 1.5	39.3 ± 2.0	54.7 ± 2.4
	MCV fL	75.5 ± 5.0	85.0 ± 5.0	93.0 ± 5.0
	MCH pg	25.5 ± 2.5	29.2 ± 2.5	32.5 ± 2.5
	MCHC g/dL	33.8 ± 3.0	34.4 ± 3.0	34.9 ± 3.0
	PLT 10 ³ /uL	62 ± 20	256 ± 40	530 ± 60

For technical assistance in the USA and Canada call Technical Service at (800) 523-3395.

All brands and products are trademarks or registered trademarks of their respective companies.