

CBC-CAL PLUS

for MINDRAY
HEMATOLOGY CALIBRATOR**CALIBRATOR****LOT PLUS115****2015-12-05**

Instrument	Parameter	Assigned Value	Acceptable Range
MINDRAY BC-5300Vet, BC-5100Vet	WBC x 10 ³ /μL	9.40	9.20 - 9.60
	RBC x 10 ⁶ /μL	4.53	4.45 - 4.61
	HGB g/dL	13.7	13.5 - 13.9
	MCV fL	94.5	92.5 - 96.5
	PLT x 10 ³ /μL	241	229 - 253
MINDRAY BC-2800Vet, BC-2600Vet (Software version 2.0 or higher)	WBC x 10 ³ /μL	9.2	9.0 - 9.4
	RBC x 10 ⁶ /μL	4.53	4.45 - 4.61
	HGB g/dL	13.8	13.6 - 14.0
	MCV fL	90.5	88.5 - 92.5
	PLT x 10 ³ /μL	246	234 - 258

NOTES

1. Before using CBC-CAL PLUS, read the package insert. Calibration errors may result if instructions are not followed exactly.
2. For technical assistance in the USA and Canada call our Technical Service Department at (800) 523-3395.

R&D Systems, Inc.614 McKinley Place NE
Minneapolis, MN USA 55413

AS146-001 Rev. 09/12