



CALIBRATOR FOR AUTOMATIC HEMATOLOGY ANALYZER(3-part)

BCC-3600

LOT : PLUS1224

Exp.date : 2025-01-05

parameter		target value
WBC	$10^9/L$	8.6
RBC	$10^{12}/L$	4.49
HGB	g/L	132
MCV	fl	82.0
PLT	$10^9/L$	224

BCC-3900

LOT : PLUS1224

Exp.date : 2025-01-05

parameter		target value
WBC	$10^9/L$	8.9
RBC	$10^{12}/L$	4.50
HGB	g/L	132
MCV	fl	81.5
PLT	$10^9/L$	236

DIRUI INDUSTRIAL CO.,LTD.

95 Yunhe Street, New & High Tech.

Development Zone Changchun, Jilin 130012, P. R. China

Tel: +86 (431) 85100409 Fax: +86 (431) 85172581

E-mail: dirui@dirui.com.cn Http://www.dirui.com.cn

AS727-000 Rev. 11/21