



## CALIBRATOR FOR AUTOMATIC HEMATOLOGY ANALYZER(5-part)

**BF-6500/6800/6880**

LOT : PLUS1224

Exp.date : 2025-01-05

parameter		target value
WBC	$10^9/L$	9.0
RBC	$10^{12}/L$	4.40
HGB	g/L	132
MCV	fl	84.0
PLT	$10^9/L$	213

**BF-6900**

LOT : PLUS1224

Exp.date : 2025-01-05

parameter		target value
WBC	$10^9/L$	8.9
RBC	$10^{12}/L$	4.51
HGB	g/L	130
MCV	fl	79.5
PLT	$10^9/L$	244

DIRUI INDUSTRIAL CO.,LTD.

95 Yunhe Street, New & High Tech.

Development Zone Changchun, Jilin 130012, P. R. China

Tel: +86 (431) 85100409 Fax: +86 (431) 85172581

E-mail: dirui@dirui.com.cn Http://www.dirui.com.cn