


CBC-CAL


HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

CALIBRATOR

LOT	PLUS1124
-----	----------

	2024-10-07
---	------------

	2024-12-05
---	------------

Applicable Instruments	Parameter	Unit	Reference Value
DH31, DH33, DH36 (Technical File Version A5.0 or higher) DH31 Vet, DH36 Vet (Technical File Version A1.0 to A4.0)	WBC	$\times 10^9/L$	9.20
	RBC	$\times 10^{12}/L$	4.53
	HGB	g/L	131
	MCV	fL	91.6
	PLT	$\times 10^9/L$	234
DH21, DH23 DH25, DH26 (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.06
	RBC	$\times 10^{12}/L$	4.50
	HGB	g/L	132
	MCV	fL	85.2
	PLT	$\times 10^9/L$	237
DH20, DH22 (Technical File Version B1.0 or higher)	WBC	$\times 10^9/L$	9.35
	RBC	$\times 10^{12}/L$	4.53
	HGB	g/L	133
	MCV	fL	89.4
	PLT	$\times 10^9/L$	244
DH31 Vet, DH36 Vet (Technical File Version A5.0 or higher)	WBC	$\times 10^9/L$	9.08
	RBC	$\times 10^{12}/L$	4.49
	HGB	g/L	128
	MCV	fL	92.9
	PLT	$\times 10^9/L$	227

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.


CBC-CAL


HEMATOLOGY CALIBRATOR

Reference Values provided by DYMIND

CALIBRATOR

LOT	PLUS1124
-----	----------

	2024-10-07
---	------------

	2024-12-05
---	------------

Applicable Instruments	Parameter	Unit	Reference Value
DH36X (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.33
	RBC	$\times 10^{12}/L$	4.48
	HGB	g/L	130
	MCV	fL	87.8
	PLT	$\times 10^9/L$	235
DP-H10 Vet (Technical File Version A1.0 or higher)	WBC	$\times 10^9/L$	9.08
	RBC	$\times 10^{12}/L$	4.49
	HGB	g/L	138
	MCV	fL	92.9
	PLT	$\times 10^9/L$	227

【NOTE】

1. The calibrator should be stored in refrigerator (2°C~8°C). After opening, it will keep stable for 7 days when it is stored airtight at 2°C~8°C.
2. Before mixing and running the calibrator after take it out from the refrigerator, please keep it at least 15 minutes until reaching room temperature(15°C~30°C).
3. Calibrator must be well mixed before using. Please mix gently, to avoid cells rupture and/or generating bubbles.
4. After using, put the calibrator back into the refrigerator to prevent contamination and evaporation.