


CBC-CAL

for E-LAB
HEMATOLOGY CALIBRATOR

CALIBRATOR

LOT PLUS0325

 2025-04-05

| Instrument | Parameter | Assigned Value | Acceptable Range |
|--|---------------|----------------|------------------|
| E-LAB EC-20、EC-21 EC-28、EC-30 EC-31、EC-36 EC-38A、EC-38B EC-38、EC-38 CRP | WBC $10^3/uL$ | 9.6 | 9.4-9.8 |
| | RBC $10^6/uL$ | 4.49 | 4.39-4.59 |
| | HGB g/dL | 13.5 | 13.3-13.7 |
| | MCV fL | 86.9 | 84.9-88.9 |
| | PLT $10^3/uL$ | 243 | 228-258 |