

# CBC-CAL PLUS

for Genrui  
HEMATOLOGY CALIBRATOR

**CALIBRATOR**

Assay values and ranges provided by Genrui

**LOT** PLUS1024



2024/11/05

Instrument	Parameter	Assigned Value	Acceptable Range
Genrui KT-6600 series,KT-6500 series, KT-8000 series,KT-6500P series, KT-6600P series,KT-8000P series	WBC x 10 <sup>9</sup> /L	9.30	9.00 - 9.60
	RBC x 10 <sup>12</sup> /L	4.26	4.11 - 4.41
	HGB g/L	126	124 - 128
	MCV fL	87.2	85.2 - 89.2
	PLT x 10 <sup>9</sup> /L	231	211 - 251
Genrui KT-6610 VET,VH50	WBC x 10 <sup>9</sup> /L	9.02	8.72 - 9.32
	RBC x 10 <sup>12</sup> /L	4.02	3.87 - 4.17
	HGB g/L	117	115 - 119
	MCV fL	91.7	89.7 - 93.7
	PLT x 10 <sup>9</sup> /L	250	230 - 270
Genrui KT-6400,KT-6300 version 2.87	WBC x 10 <sup>9</sup> /L	9.5	9.2 - 9.8
	RBC x 10 <sup>12</sup> /L	4.30	4.15 - 4.45
	HGB g/L	129	127 - 131
	MCV fL	88.7	86.7 - 90.7
	PLT x 10 <sup>9</sup> /L	250	230 - 270
Genrui New KT-6300,KT-6390,KT-6400	WBC x 10 <sup>9</sup> /L	9.73	9.43 - 10.03
	RBC x 10 <sup>12</sup> /L	4.56	4.41 - 4.71
	HGB g/L	129	127 - 131
	MCV fL	88.9	86.9 - 90.9
	PLT x 10 <sup>9</sup> /L	229	209 - 249
Genrui KT-60,KT-40	WBC x 10 <sup>9</sup> /L	9.66	9.36 - 9.96
	RBC x 10 <sup>12</sup> /L	4.50	4.35 - 4.65
	HGB g/L	129	127 - 131
	MCV fL	87.7	85.7 - 89.7
	PLT x 10 <sup>9</sup> /L	225	205 - 245
Genrui KT-6300 VET	WBC x 10 <sup>9</sup> /L	8.7	8.2 - 9.2
	RBC x 10 <sup>12</sup> /L	4.50	4.20 - 4.80
	HGB g/L	126	121 - 131
	MCV fL	90.0	87.0 - 93.0
	PLT x 10 <sup>9</sup> /L	213	183 - 243
Genrui VH 20,30,40	WBC x 10 <sup>9</sup> /L	7.90	7.40 - 8.40
	RBC x 10 <sup>12</sup> /L	4.15	3.85 - 4.45
	HGB g/L	112	107 - 117
	MCV fL	90.7	87.7 - 93.7
	PLT x 10 <sup>9</sup> /L	231	201 - 261

- Notes:**
1. For expiry date please refer to the label on tube. Once opened, please use within 1 week if kept at 2-8°C.
  2. Before using, please gently mix well.
  3. All brands and products are trademarks or registered trademarks of their respective companies.

## Genrui Biotech Inc.

Address:4-10F, Building 3, Geya Technology Park, Guangming District,  
518106, Shenzhen, P.R.China  
Tel: +86 755 26835560 ext: 8026 Fax: +86 755 26678789