

CBC-CAL PLUS

for HD Consortium India Limited
HEMATOLOGY CALIBRATOR

Assay values and ranges provided by HD Consortium India Limited

CALIBRATOR

LOT PLUS0224



2024/03/05

Instrument	Parameter	Assigned Value	Acceptable Range
HDCount 5, HDCount Advance	WBC x 10 ⁹ /L	9.26	8.96 - 9.56
	RBC x 10 ¹² /L	4.44	4.29 - 4.59
	HGB g/L	131	129 - 133
	MCV fL	89.3	87.3 - 91.3
	PLT x 10 ⁹ /L	239	219 - 259
HDCount 3	WBC x 10 ⁹ /L	9.76	9.46 - 10.06
	RBC x 10 ¹² /L	4.64	4.49 - 4.79
	HGB g/L	134	132 - 136
	MCV fL	91.1	89.1 - 93.1
	PLT x 10 ⁹ /L	231	211 - 251

- Notes:**
1. For expiry date please refer to the label on tube. Once opened, please use within 1 week if kept at 2-8°C.
 2. Before using, please gently mix well.
 3. All brands and products are trademarks or registered trademarks of their respective companies.

HD Consortium India Limited

Address: A-11, Okhla Industrial Area, Phase-1, South Delhi, Delhi, 110020

Web: www.hdcindia.com Tel: +91 98186 36254