


# CBC-CAL PLUS

for Linear  
HEMATOLOGY CALIBRATOR

**CALIBRATOR**

Assay values and ranges provided by Linear

**LOT** PLUS0124  
 2024/02/05

Instrument	Parameter	Assigned Value	Acceptable Range
Linear STEL 5,STEL 5 PLUS	WBC x 10 <sup>9</sup> /L	9.54	9.24 - 9.84
	RBC x 10 <sup>12</sup> /L	4.32	4.17 - 4.47
	HGB g/L	130	128 - 132
	MCV fL	90.0	88.0 - 92.0
	PLT x 10 <sup>9</sup> /L	243	223 - 263

- Notes:**
1. For expiry date please refer to the label on tube. Once opened, please use within 1 week if kept at 2-8°C.
  2. Before using, please gently mix well.
  3. All brands and products are trademarks or registered trademarks of their respective companies.

**LINEAR**

LINEAR CHEMICALS, S.L.U. Joaquim Costa 18 2ª planta. 08390 Montgat (Barcelona) SPAIN  
Telf. (+34) 934 694 990; E-mail: info@linear.es ; website: www.linear.es NIF-VAT:B60485687