

HEMATOLOGY CALIBRATOR

ASSAY VALUES AND EXPECTED RANGES

MODEL: LW-5CaI LOT: LW23001 MFG: 2022.11.07

EXP: 2023.02.05

Applies	Parameter	SI	Mean	Range
LWD6580、 LWD6680	WBC	$\times 10^9/L$	8.55	± 0.30
	RBC	$\times 10^{12}/L$	4.51	± 0.10
	HGB	g/L	133	± 2
	MCV	fL	90.9	± 2.0
	PLT	$\times 10^9/L$	295	± 15

PS: The results applied only to the reagents of Landwind.

HEMATOLOGY CALIBRATOR

ASSAY VALUES AND EXPECTED RANGES

MODEL: LW-5Cal LOT: LW23001 MFG: 2022.11.07

EXP: 2023.02.05

Applies	Parameter	SI	Mean	Range
LWD6500、 LWD6600	WBC	$\times 10^9/L$	8.47	± 0.30
	RBC	$\times 10^{12}/L$	4.31	± 0.10
	HGB	g/L	137	± 2
	MCV	fL	90.8	± 2.0
	PLT	$\times 10^9/L$	260	± 15

PS: The results applied only to the reagents of Landwind.