

CBC-CAL PLUS

For Zybio

HEMATOLOGY CALIBRATOR

CALIBRATOR

Assay values and ranges provided by Zybio

LOT PLUS1124

 reference tag

Instrument Model	Parameter	Target Value	Acceptable Range
Z3,Z3 CRP	WBC $\times 10^9/L$	9.22	± 0.27
	RBC $\times 10^{12}/L$	4.35	± 0.12
Z30,Z30 CRP	HGB g/L	131.0	± 2.2
	MCV fL	89.9	± 2.4
Z31,Z31 CRP	PLT $\times 10^9/L$	243	± 17
Z5,Z51	WBC $\times 10^9/L$	8.82	± 0.27
	RBC $\times 10^{12}/L$	4.54	± 0.12
	HGB g/L	128.0	± 2.3
	MCV fL	85.1	± 2.4
	PLT $\times 10^9/L$	236	± 17
Z50,Z52	WBC $\times 10^9/L$	9.60	± 0.28
	RBC $\times 10^{12}/L$	4.55	± 0.12
	HGB g/L	129.0	± 2.2
	MCV fL	85.2	± 2.4
	PLT $\times 10^9/L$	240	± 17
EXZ 6000,EXZ 6010,EXZ 6020 EXZ 6000Pro,EXZ 6010Pro, EXZ 6020Pro	WBC $\times 10^9/L$	9.93	± 0.29
	RBC $\times 10^{12}/L$	4.29	± 0.12
	HGB g/L	127.0	± 2.2
	MCV fL	84.4	± 2.3
	PLT $\times 10^9/L$	254	± 18

Notes: Before using CBC-CAL PLUS, read the package insert. Calibration errors may result if instructions are not followed exactly.

Zybio Inc.

Add: Floor 1 to Floor 5, Building 30, No.6 of Taikang Road, Block C of Jianqiao Industrial Park, Dadukou District, Chongqing, China 400082 Tel: +86(0)23 6865 9999 Email: info@zybio.com Web: www.zybio.com