

Tech-Cal

HEMATOLOGY CALIBRATOR

CALIBRATOR



LOT	T1224
------------	-------



2025-01-05

Calibration Values for ADVIA® 120, 2120 & 2120i Hematology Systems

Parameter	Calibration Values		
WBCB × 10 ³ /μL (BASO)	7.15		
WBCP × 10 ³ /μL (PEROX)	7.50		
RBC × 10 ⁶ /μL	4.43		
HGB g/dL	13.4		
MCV fL	78.5		
CHCM %	40.0		
PLT × 10 ³ /μL	207		
NEUTx %	54.5	± 6.0	
NEUTy %	75.0	± 5.0	

Before using, refer to the instruction sheet for mixing directions.

For technical assistance in the USA and Canada call Technical Service at (800) 523-3395.

All brands and products are trademarks or registered trademarks of their respective companies.