


# CBC-DH




For Rayto

HEMATOLOGY CONTROLS

**CONTROL**

 DH2501

 2025.03.05

Instrument	Parameters	Units	Low  DH2501L		Normal  DH2501N		High  DH2501H	
			Mean	Range	Mean	Range	Mean	Range
Hemaray 50CRP	WBC	$\times 10^9/L$	3.6	$\pm 0.5$	8.3	$\pm 1.0$	18.9	$\pm 2.5$
Hemaray 51								
Hemaray 52	NEU	%	46.1	$\pm 15$	53.2	$\pm 15$	60.8	$\pm 10$
Hemaray 53								
Hemaray 54	LYM	%	38.6	$\pm 15$	26.6	$\pm 15$	18.0	$\pm 10$
Hemaray 55CRP								
Hemaray 56CRP	MON	%	7.2	$\pm 5$	8.2	$\pm 5$	8.0	$\pm 5$
Hemaray 57CRP								
Hemaray 58CRP	EOS	%	7.2	$\pm 5$	11.2	$\pm 5$	12.4	$\pm 5$
Hemaray 59CRP								
	BAS	%	0.9	$\pm 0.9$	0.8	$\pm 0.8$	0.8	$\pm 0.8$
	NEU	$\times 10^9/L$	1.66	$\pm 0.5$	4.42	$\pm 1.5$	11.49	$\pm 2.8$
	LYM	$\times 10^9/L$	1.37	$\pm 0.5$	2.20	$\pm 1.5$	3.40	$\pm 2.8$
	MON	$\times 10^9/L$	0.26	$\pm 0.2$	0.68	$\pm 0.5$	1.51	$\pm 1.0$
	EOS	$\times 10^9/L$	0.26	$\pm 0.2$	0.93	$\pm 0.50$	2.34	$\pm 1.00$
	BAS	$\times 10^9/L$	0.03	$\pm 0.03$	0.06	$\pm 0.06$	0.13	$\pm 0.13$
	RBC	$\times 10^{12}/L$	2.42	$\pm 0.18$	4.77	$\pm 0.24$	5.55	$\pm 0.30$
	HGB	g/L	62	$\pm 4$	135	$\pm 6$	170	$\pm 8$
	MCHC	g/L	335	$\pm 40$	337	$\pm 40$	340	$\pm 40$
	MCH	pg	25.6	$\pm 3.0$	28.3	$\pm 3.0$	30.6	$\pm 3.0$
	MCV	fL	76.6	$\pm 5$	84.0	$\pm 5$	90.0	$\pm 5$
	RDW-CV	%	13.1	$\pm 3.0$	13.6	$\pm 3.0$	13.6	$\pm 3.0$
	HCT	%	18.5	$\pm 2.5$	40.1	$\pm 3.5$	50.0	$\pm 4.5$
	PLT	$\times 10^9/L$	51	$\pm 20$	235	$\pm 40$	503	$\pm 60$
	MPV	fL	9.7	$\pm 3.0$	9.3	$\pm 3.0$	9.8	$\pm 3.0$