


# CBC-DH




For Rayto

HEMATOLOGY CONTROLS

**CONTROL**

 DH2501

 2025.03.05

Instrument	Parameters	Units	Low  DH2501L		Normal  DH2501N		High  DH2501H	
			Mean	Range	Mean	Range	Mean	Range
Hemaray 83	WBC	$\times 10^9/L$	3.4	$\pm 0.5$	8.4	$\pm 1.0$	20.5	$\pm 2.5$
Hemaray 86	NEU	%	48.5	$\pm 15$	52.3	$\pm 15$	62.4	$\pm 10$
Hemaray 89	LYM	%	39.7	$\pm 15$	28.8	$\pm 15$	19.3	$\pm 10$
	MON	%	4.3	$\pm 4.3$	6.5	$\pm 5$	6.3	$\pm 5$
	EOS	%	6.7	$\pm 5$	11.6	$\pm 5$	11.2	$\pm 5$
	BAS	%	0.8	$\pm 0.8$	0.8	$\pm 0.8$	0.8	$\pm 0.8$
	NEU	$\times 10^9/L$	1.65	$\pm 0.5$	4.39	$\pm 1.5$	12.79	$\pm 2.8$
	LYM	$\times 10^9/L$	1.36	$\pm 0.5$	2.41	$\pm 1.5$	3.96	$\pm 2.8$
	MON	$\times 10^9/L$	0.15	$\pm 0.15$	0.54	$\pm 0.5$	1.28	$\pm 1.0$
	EOS	$\times 10^9/L$	0.23	$\pm 0.2$	0.97	$\pm 0.50$	2.29	$\pm 1.00$
	BAS	$\times 10^9/L$	0.01	$\pm 0.01$	0.07	$\pm 0.07$	0.16	$\pm 0.16$
	RBC	$\times 10^{12}/L$	2.26	$\pm 0.18$	4.41	$\pm 0.24$	5.19	$\pm 0.30$
	HGB	g/L	61	$\pm 4$	132	$\pm 6$	167	$\pm 8$
	MCHC	g/L	339	$\pm 40$	344	$\pm 40$	346	$\pm 40$
	MCH	pg	27.0	$\pm 3.0$	29.9	$\pm 3.0$	32.2	$\pm 3.0$
	MCV	fL	79.6	$\pm 5$	87.0	$\pm 5$	92.9	$\pm 5$
	RDW-CV	%	16.8	$\pm 3.0$	16.2	$\pm 3.0$	15.6	$\pm 3.0$
	HCT	%	18.0	$\pm 2.5$	38.4	$\pm 3.5$	48.2	$\pm 4.5$
	PLT	$\times 10^9/L$	48	$\pm 20$	245	$\pm 40$	581	$\pm 60$
	MPV	fL	8.1	$\pm 3.0$	8.1	$\pm 3.0$	8.4	$\pm 3.0$