

CBC-DH

For Rayto

HEMATOLOGY CONTROLS

CONTROL

LOT

 DH2609

2026.11.05

		Low		Normal		High		
		LOT	DH2609L	LOT	DH2609N	LOT	DH2609H	
Instrument	Parameters	Units	Mean	Range	Mean	Range	Mean	Range
Hemaray 90 Hemaray 91 Hemaray 92	WBC	×10 ⁹ /L	3.61	±0.50	9.04	±1.00	21.00	±2.50
	NEU	%	46.8	±15.0	55.6	±15.0	64.4	±10.0
	LYM	%	40.8	±15.0	30.5	±15.0	21.3	±10.0
	MON	%	7.1	±5.0	7.9	±5.0	7.6	±5.0
	EOS	%	4.5	±4.5	5.2	±5.0	5.9	±5.0
	BAS	%	0.8	±0.8	0.8	±0.8	0.8	±0.8
	NEU	×10 ⁹ /L	1.69	±0.50	5.03	±1.50	13.52	±2.80
	LYM	×10 ⁹ /L	1.47	±0.50	2.76	±1.50	4.47	±2.80
	MON	×10 ⁹ /L	0.26	±0.20	0.71	±0.50	1.60	±1.00
	EOS	×10 ⁹ /L	0.16	±0.16	0.47	±0.47	1.24	±1.00
	BAS	×10 ⁹ /L	0.03	±0.03	0.07	±0.07	0.17	±0.17
	RBC	×10 ¹² /L	2.33	±0.18	4.45	±0.24	5.37	±0.30
	HGB	g/L	61	±4	133	±6	169	±8
	MCHC	g/L	328	±40	338	±40	342	±40
	MCH	pg	26.2	±3.0	29.9	±3.0	31.5	±3.0
	MCV	fL	79.9	±5.0	88.4	±5.0	91.9	±5.0
	RDW-CV	%	19.6	±3.0	17.3	±3.0	17.7	±3.0
	RDW-SD	fL	68.8	±10.0	68.2	±10.0	73.0	±12.0
	HCT	%	18.6	±2.5	39.3	±3.5	49.4	±4.5
	PLT	×10 ⁹ /L	53	±20	256	±40	508	±60
	MPV	fL	8.8	±3.0	9.9	±3.0	9.9	±3.0
	PDW	/	11.2	±3.0	12.3	±3.0	12.4	±3.0
	PCT	%	0.047	±0.047	0.253	±0.100	0.503	±0.200
	P-LCR	%	15.1	±8.0	25.0	±8.0	24.6	±8.0
	P-LCC	×10 ⁹ /L	8	±8	64	±25	125	±35