

**CBC-3D<sup>®</sup>**

For Aehealth

Hematology Controls.

**LOT** B1122



2023-2-5

**CONTROL**

**ASSAY VALUES AND EXPECTED RANGES**

Instrument	Parameter	Low	Normal	High
	WBC 10 <sup>9</sup> /L	2.2±0.5	7.8±1.0	20.9±2.5
	RBC 10 <sup>12</sup> /L	2.3±0.2	4.59±0.24	5.75±0.3
	HGB g/L	59.0±4	136.4±6	189.0±8
	HCT %	17.1±1.5	43.36±2.0	55.8±2.4
	MCV fL	78.2±5.0	94.48±5.0	99.2±5.0
AerC-3	PLT 10 <sup>9</sup> /L	80.0±20	254.6±40	578±60
	LYM %	54.0±12.0	28.7±10.0	13.5±6.0
	MONO %	15.0±10.0	9.4±8.0	5.8±5.0
	GR %	31.0±9.0	61.9±9.0	81±8.0
	LYM 10 <sup>9</sup> /L	1.2±0.3	2.2±0.7	3.1±1.4
	MONO 10 <sup>9</sup> /L	0.3±0.2	0.7±0.6	1.3±1.1
	GR 10 <sup>9</sup> /L	0.7±0.2	4.7±0.7	17.5±1.8

Before using, refer to the instruction sheet for mixing directions.