

# CBC-CAL PLUS


## HEMATOLOGY CALIBRATOR

**CALIBRATOR**

---

 Assay values and expected ranges

LOT PLUS0323

  
 2023/05/05

Method	Parameter	Calibration Values	Acceptable Range
hemaline 500 V4.01.160421 (WIC- FPGA)	WBC 10 <sup>9</sup> /L	9.56	9.36 - 9.76
	WOC 10 <sup>9</sup> /L	8.77	8.57 - 8.97
	RBC 10 <sup>12</sup> /L	4.49	4.44 - 4.54
	HGB g/dL	13.4	13.2 - 13.6
	MCV fL	83.0	82.0 - 84.0
	PLT 10 <sup>9</sup> /L	259	249 - 269
hemaline 300 application software V4.64.170701 And above (Including V4.64)	WBC 10 <sup>9</sup> /L	9.7	9.5 - 9.9
	RBC 10 <sup>12</sup> /L	4.70	4.65 - 4.75
	HGB g/dL	13.6	13.4 - 13.8
	MCV fL	93.1	92.1 - 94.1
	PLT 10 <sup>9</sup> /L	248	238 - 258

Before using, refer to the instruction sheet for mixing directions.

**MDX MEDTEK DIAGNOSTICS (NETHERLANDS) B.V.**  
 Jollemanhof 148, 1019GW Amsterdam, The Netherlands