

XERet Control

HEMATOLOGY CONTROLS

ASSAY VALUES AND EXPECTED RANGES

CONTROL

maccura®F810 (manual sampling)

QCP DATA MONTHS: JANUARY, FEBRUARY

 2023-03-05

Parameters	LEVEL 1 LOT 33011001			LEVEL 2 LOT 33011002			LEVEL 3 LOT 33011003		
	Assay Mean	±Limit	Expected Range	Assay Mean	±Limit	Expected Range	Assay Mean	±Limit	Expected Range
RET# (10 ⁹ /L)	4.5±1.4		3.1~5.9	226.1±67.8		158.3~293.9	286.3±85.9		200.4~372.2
RET% (%)	0.10±0.10		0.00~0.20	4.69±1.41		3.28~6.10	11.05±3.32		7.73~14.37
HFR (%)	2.2±6.6		0.0~8.8	5.7±11.5		0.0~17.2	4.7±23.5		0.0~28.2
MFR (%)	7.7±15.4		0.0~23.1	15.2±15.2		0.0~30.4	21.4±32.1		0.0~53.5
LFR (%)	90.1±45.1		45.0~135.2	79.1±39.6		39.5~118.7	73.9±37.0		36.9~110.9
IRF% (%)	9.9±19.8		0.0~29.7	20.9±20.9		0.0~41.8	26.1±52.2		0.0~78.3
RET-He (pg)	18.5±1.9		16.6~20.4	15.6±1.6		14.0~17.2	14.1±1.4		12.7~15.5
RBC-O (10 ¹² /L)	5.40±0.54		4.86~5.94	4.39±0.44		3.95~4.83	2.40±0.24		2.16~2.64

- NOTES:**
1. Before using, confirm the software for the latest version, refer to the instruction sheet for mixing directions, and choice the QC mode.
 2. All brands and products are trademarks or registered trademarks of their respective companies, All results are for reference only.

XERet Control

HEMATOLOGY CONTROLS

ASSAY VALUES AND EXPECTED RANGES

CONTROL

maccura®F880 (manual sampling)

QCP DATA MONTHS: JANUARY, FEBRUARY



2023-03-05

Parameters	LEVEL 1 LOT 33011001			LEVEL 2 LOT 33011002			LEVEL 3 LOT 33011003		
	Assay Mean	±Limit	Expected Range	Assay Mean	±Limit	Expected Range	Assay Mean	±Limit	Expected Range
RET# (10 ⁹ /L)	4.5±1.4		3.1~5.9	226.1±67.8		158.3~293.9	286.3±85.9		200.4~372.2
RET% (%)	0.10±0.10		0.00~0.20	4.69±1.41		3.28~6.10	11.05±3.32		7.73~14.37
HFR (%)	2.2±6.6		0.0~8.8	5.7±11.5		0.0~17.2	4.7±23.5		0.0~28.2
MFR (%)	7.7±15.4		0.0~23.1	15.2±15.2		0.0~30.4	21.4±32.1		0.0~53.5
LFR (%)	90.1±45.1		45.0~135.2	79.1±39.6		39.5~118.7	73.9±37.0		36.9~110.9
IRF% (%)	9.9±19.8		0.0~29.7	20.9±20.9		0.0~41.8	26.1±52.2		0.0~78.3
RET-He (pg)	18.5±1.9		16.6~20.4	15.6±1.6		14.0~17.2	14.1±1.4		12.7~15.5
RBC-O (10 ¹² /L)	5.40±0.54		4.86~5.94	4.39±0.44		3.95~4.83	2.40±0.24		2.16~2.64

- NOTES:**
1. Before using, confirm the software for the latest version, refer to the instruction sheet for mixing directions, and choice the QC mode.
 2. All brands and products are trademarks or registered trademarks of their respective companies, All results are for reference only.